

Gascontrol center

GASCONTROL CENTER

Produkte Benutzer Messdaten Extra ?

MAPY 4.0 192.168.1.238

	PROD_NAME	PROD_NO	BARCODE	CHARGE	USER	MEAS.TIME	TIME	DATE	INDEX	INFO	GAS_O2	GAS_CO2	GAS_Rest	AL_O2min	AL_O2max	AL_CO2min	AL_CO2max
0	Salami	678126	4007817304679	1	ADMINISTRATOR	10 14:48:29	16.05.2011	350	OK		0,23	24,8	75,0	0,0	0,5	24,0	30,0
1	Salami	678126	4007817304679	1	ADMINISTRATOR	10 14:48:22	16.05.2011	351	OK		0,19	25,1	74,7	0,0	0,5	24,0	30,0
2	Salami	678126	4007817304679	1	Miller	10 14:50:37	16.05.2011	352	OK		0,27	25,2	74,5	0,0	0,5	24,0	30,0
3	Salami	678126	4007817304679	1	Miller	10 14:51:53	16.05.2011	353	OK		0,17	24,9	74,9	0,0	0,5	24,0	30,0
4	Salami	678126	4007817304679	1	Miller	10 14:53:57	16.05.2011	354	OK		0,22	24,7	75,1	0,0	0,5	24,0	30,0
5	Salami	678126	4007817304679	1	Miller	10 14:54:48	16.05.2011	355	OK		0,26	25,0	74,7	0,0	0,5	24,0	30,0
6	Salami	678126	4007817304679	1	Miller	10 14:55:39	16.05.2011	356	OK		0,11	25,2	74,7	0,0	0,5	24,0	30,0
7	Salami	678126	4007817304679	1	Miller	10 14:56:31	16.05.2011	357	OK		0,13	24,9	75,0	0,0	0,5	24,0	30,0
8	Salami	678126	4007817304679	1	Miller	10 14:57:22	16.05.2011	358	FAIL O2		0,56	24,1	75,3	0,0	0,5	24,0	30,0
9	Salami	678126	4007817304679	1	Miller	10 14:58:13	16.05.2011	359	OK		0,23	25,2	74,6	0,0	0,5	24,0	30,0
10	Mozzarella	776634	4007817304457	5	Berger	10 15:00:47	16.05.2011	360	OK		0,29	29,7	70,0	0,0	0,5	28,0	32,0
11	Mozzarella	776634	4007817304457	5	Berger	10 15:01:36	16.05.2011	361	OK		0,24	30,1	69,7	0,0	0,5	28,0	32,0
12	Mozzarella	776634	4007817304457	5	Berger	10 15:02:30	16.05.2011	362	OK		0,18	30,0	69,8	0,0	0,5	28,0	32,0
13	Mozzarella	776634	4007817304457	5	Berger	10 15:03:21	16.05.2011	363	OK		0,22	29,9	69,9	0,0	0,5	28,0	32,0
14	Mozzarella	776634	4007817304457	5	Berger	10 15:04:12	16.05.2011	364	OK		0,26	30,2	69,5	0,0	0,5	28,0	32,0
15	Mozzarella	776634	4007817304457	5	Berger	10 15:05:04	16.05.2011	365	OK		0,29	29,7	70,0	0,0	0,5	28,0	32,0
16	Mozzarella	776634	4007817304457	5	Berger	10 15:05:55	16.05.2011	366	OK		0,19	29,8	70,0	0,0	0,5	28,0	32,0
17	Mozzarella	776634	4007817304457	5	Berger	10 15:06:46	16.05.2011	367	OK		0,23	30,1	69,7	0,0	0,5	28,0	32,0
18	Mozzarella	776634	4007817304457	5	Berger	10 15:07:38	16.05.2011	368	OK		0,22	30,0	69,8	0,0	0,5	28,0	32,0
19	Mozzarella	776634	4007817304457	5	Berger	10 15:08:03	16.05.2011	368	OK		0,24	29,7	70,1	0,0	0,5	28,0	32,0
20	Mortadella	534678	4007817304311	3	Kirst	12 15:08:03	16.05.2011	368	OK		0,30	28,0	71,7	0,0	0,5	26,0	30,0
21	Mortadella	534678	4007817304311	3	Kirst	12 15:08:31	16.05.2011	370	OK		0,27	27,9	71,8	0,0	0,5	26,0	30,0
22	Mortadella	534678	4007817304311	3	Kirst	12 15:08:52	16.05.2011	371	OK		0,26	28,1	71,6	0,0	0,5	26,0	30,0
23	Mortadella	534678	4007817304311	3	Kirst	12 15:09:13	16.05.2011	372	FAIL CO2		0,48	22,2	77,3	0,0	0,5	26,0	30,0
24	Mortadella	534678	4007817304311	3	Kirst	12 15:09:36	16.05.2011	373	OK		0,32	27,8	71,9	0,0	0,5	26,0	30,0
25	Mortadella	534678	4007817304311	3	Kirst	12 15:09:57	16.05.2011	374	OK		0,19	28,1	71,7	0,0	0,5	26,0	30,0
26	Mortadella	534678	4007817304311	3	Kirst	12 15:10:24	16.05.2011	375	OK		0,22	27,7	72,1	0,0	0,5	26,0	30,0
27	Mortadella	534678	4007817304311	3	Kirst	12 15:10:41	16.05.2011	376	OK		0,18	28,0	71,8	0,0	0,5	26,0	30,0
28	Mortadella	534678	4007817304311	3	Kirst	12 15:11:00	16.05.2011	377	OK		0,21	27,9	71,9	0,0	0,5	26,0	30,0
29	Mortadella	534678	4007817304311	3	Kirst	12 15:11:25	16.05.2011	378	OK		0,24	27,9	71,9	0,0	0,5	26,0	30,0
30																	

GASCONTROL 

Windows-Software for the documentation of analysis-results of MAPY, LEAK-MASTER®, MAPMAX and other equipment

with integrated or separate analysis.

Benefits

- quality control by means of data transfer and storing
 - date and time
 - user (name or examiner)
 - lot data
 - product line data
- export data interface for further interpretation of measured data in MS EXCEL®
 - data transfer to Control Unit (touch screen control)
 - product data
- administration of product names
- administration of user names inclusive access code
- logging and readable from up to 21 analysers
- multilingual interface: German, English, Spanish, Italian, Polish, French, Dutch (more to follow)

System requirements

Operating system	WINDOWS® 2000, XP, Vista, 7
Hardware	graphic resolution min. 800 x 600 pixel standard Ethernet cable installed by customer
Compatibility	all MAPY, LEAK-MASTER®, MAPMAX and gas mixers with integrated analysers - please specify device number with order -